

## Deutsche Akkreditierungsstelle GmbH

### Annex to the Accreditation Certificate D-K-17723-01-00 according to DIN EN ISO/IEC 17025:2018

**Valid from:** 08.07.2021

**Date of issue** 08.07.2021

Holder of certificate:

**WENZEL Metrology GmbH**  
**Werner-Wenzel-Straße, 97859 Wiesthal**

Calibration in the fields:

**Dimensional quantities**

**Coordinate measuring technology**

— **Coordinate measuring machines <sup>a)</sup>**

<sup>a)</sup> On-Site Calibration

*The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories. Laboratories that conform to the requirements of this standard, operate generally in accordance with the principles of DIN EN ISO 9001.*

*The certificate together with the annex reflects the status as indicated by the date of issue.*

*The current status of any given scope of accreditation may be found respectively in the database of accredited bodies of Deutsche Akkreditierungsstelle GmbH <https://www.dakks.de/en/content/accredited-bodies-dakks>.*

Annex to the accreditation certificate D-K-17723-01-00

On-site Calibration

Calibration and Measurement Capabilities (CMC)

| Measurement quantity / Calibration item                                                                                                                                                                                                                                                                                                                                                                                                | Range                                                                                              | Measurement conditions / procedure                                                                                                                                                                                             | Expanded uncertainty of measurement <sup>1)</sup>                                                                                                                                                                                                                                                                                                                                                                                                                                          | Remarks               |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------|
| <b>Coordinate measuring technology</b><br>Coordinate measuring machines with control software<br>Metrosoft CM and QUARTIS, Software from WENZEL Metromec Software AG, Chur, CH<br>Modus Metrology, Software from Renishaw plc., Wotton-under-Edge, UK<br>INCA3D, Software from Mora Metrology GmbH, Aschaffenburg, GER<br>PolyWorks, Software from Duwe-3d AG, Lindau, GER<br>Metrologic, Software from Metrologic Group SA, Meylan, F | Coordinate measuring machines featuring a measuring volume with a space diagonal of $\leq 4666$ mm | Calibration of metrological characteristics according to guideline: DKD-R 4-3 part 18.1:2018, and the following standards DIN EN ISO 10360                                                                                     |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |                       |
|                                                                                                                                                                                                                                                                                                                                                                                                                                        |                                                                                                    | Determination of the error of indication for size measurement $E_L$ ( $E_0$ and $E_{150}$ ) by using step gauges according to DIN EN ISO 10360-2:2010 (One conjunction measurement by displacement of the measuring standard.) | without temperature compensation:<br>$0,1 \mu\text{m} + 0,35 \cdot 10^{-6} \cdot l$<br><br>without temperature compensation and with one conjunction measurement:<br>$0,2 \mu\text{m} + 0,35 \cdot 10^{-6} \cdot l$<br><br>with temperature compensation:<br>$0,1 \mu\text{m} + 0,40 \cdot 10^{-6} \cdot l$<br>with $\Delta T = 2$ K<br><br>with temperature compensation and with one conjunction measurement:<br>$0,2 \mu\text{m} + 0,40 \cdot 10^{-6} \cdot l$<br>with $\Delta T = 2$ K | $l$ = measured length |
|                                                                                                                                                                                                                                                                                                                                                                                                                                        |                                                                                                    | Determination of repeatability range $R_0$ according to DIN EN ISO 10360-2:2010                                                                                                                                                | 0,07 $\mu\text{m}$                                                                                                                                                                                                                                                                                                                                                                                                                                                                         |                       |
|                                                                                                                                                                                                                                                                                                                                                                                                                                        |                                                                                                    | Determination of individual switch-form deviation $P_{\text{Form.Sph.1x25:SS:Tact}}$ by reference sphere according to DIN EN ISO 10360-5:2020                                                                                  | 0,13 $\mu\text{m}$                                                                                                                                                                                                                                                                                                                                                                                                                                                                         |                       |
|                                                                                                                                                                                                                                                                                                                                                                                                                                        |                                                                                                    | Determination of individual switch-dimensional deviation $P_{\text{Size.Sph.1x25:SS:Tact}}$ by reference sphere according to DIN EN ISO 10360-5:2020                                                                           | 0,14 $\mu\text{m}$                                                                                                                                                                                                                                                                                                                                                                                                                                                                         |                       |

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Date of issue: 08.07.2021

Valid from: 08.07.2021

Page 2 of 4

- Translation -

Annex to the accreditation certificate D-K-17723-01-00

On-site Calibration

Calibration and Measurement Capabilities (CMC)

| Measurement quantity / Calibration item                                                                                                                                                                                                                                                                                                                                                      | Range                                                                                              | Measurement conditions / procedure                                                                                                                             | Expanded uncertainty of measurement <sup>1)</sup>                                                                                                                                                                                                      | Remarks               |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------|
|                                                                                                                                                                                                                                                                                                                                                                                              |                                                                                                    | Determination of form deviation $P_{Form.Sph.Scan:PP:Tact}$ at scanning mode by reference sphere according to DIN EN ISO 10360-5:2020                          | 0,13 $\mu\text{m}$                                                                                                                                                                                                                                     |                       |
|                                                                                                                                                                                                                                                                                                                                                                                              |                                                                                                    | Determination of dimensional deviation $P_{Size.Sph.Scan:PP:Tact}$ at scanning mode by reference sphere according to DIN EN ISO 10360-5:2020                   | 0,14 $\mu\text{m}$                                                                                                                                                                                                                                     |                       |
|                                                                                                                                                                                                                                                                                                                                                                                              |                                                                                                    | Determination of time span in scanning mode $\tau_{Sph.Scan:PP:Tact}$ according to DIN EN ISO 10360-5:2020                                                     | 0,28 s                                                                                                                                                                                                                                                 |                       |
| Coordinate measuring machines with control software<br>Metrosoft CM and QUARTIS, Software from WENZEL Metromec Software AG, Chur, CH<br>Modus Metrology, Software from Renishaw plc., Wotton-under-Edge, UK<br>INCA3D, Software from Mora Metrology GmbH, Aschaffenburg, GER<br>PolyWorks, Software from Duwe-3d AG, Lindau, GER<br>Metrologic, Software from Metrologic Group SA, Meylan, F | Coordinate measuring machines featuring a measuring volume with a space diagonal of $\leq 9090$ mm | Determination of the error of indication for size measurement $E_L$ ( $E_0$ and $E_{150}$ ) by using demountable ball bar according to DIN EN ISO 10360-2:2010 | without temperature compensation:<br>$2 \cdot \sqrt{l} \cdot (0,4 \mu\text{m} + 0,55 \cdot 10^{-6} \cdot l)$<br><br>with temperature compensation:<br>$2 \cdot \sqrt{l} \cdot (0,4 \mu\text{m} + 0,61 \cdot 10^{-6} \cdot l)$<br>with $\Delta T = 2$ K | $l$ = measured length |
|                                                                                                                                                                                                                                                                                                                                                                                              |                                                                                                    | Determination of repeatability range $R_0$ according to DIN EN ISO 10360-2:2010                                                                                | 0,19 $\mu\text{m}$                                                                                                                                                                                                                                     |                       |
|                                                                                                                                                                                                                                                                                                                                                                                              |                                                                                                    | Determination of individual switch-form deviation $P_{Form.Sph.1x25:SS:Tact}$ by reference sphere according to DIN EN ISO 10360-5:2020                         | 0,13 $\mu\text{m}$                                                                                                                                                                                                                                     |                       |
|                                                                                                                                                                                                                                                                                                                                                                                              |                                                                                                    | Determination of individual switch-dimensional deviation $P_{Size.Sph.1x25:SS:Tact}$ by reference sphere according to DIN EN ISO 10360-5:2020                  | 0,14 $\mu\text{m}$                                                                                                                                                                                                                                     |                       |
|                                                                                                                                                                                                                                                                                                                                                                                              |                                                                                                    |                                                                                                                                                                |                                                                                                                                                                                                                                                        |                       |

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Date of issue: 08.07.2021

Valid from: 08.07.2021

Page 3 of 4

- Translation -

Annex to the accreditation certificate D-K-17723-01-00

On-site Calibration

Calibration and Measurement Capabilities (CMC)

| Measurement quantity /<br>Calibration item                                                                                                                                                     | Range                                                                                                         | Measurement conditions<br>/ procedure                                                                                                                | Expanded uncertainty<br>of measurement <sup>1)</sup>                                              | Remarks |
|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------|---------|
| Coordinate measuring<br>machines with CT-<br>sensor and control<br>software Control<br>according to evaluation<br>software<br>Metrosoft QUARTIS of<br>WENZEL Metromec<br>Software AG, Chur, CH | Coordinate measuring<br>machines featuring a<br>measuring volume with a<br>space diagonal of<br>$\leq 410$ mm | Calibration of metrological<br>characteristics according to<br>guideline:<br>DKD-R 4-3 part 18.1:2018<br>and the following standards<br>VDI/VDE 2630 |                                                                                                   |         |
|                                                                                                                                                                                                |                                                                                                               | Determination of error of<br>indication for size<br>measurement $E_{(TS)}$ by CT-<br>artefact according to<br>VDI/VDE 2630 part 1.3:2011             | for length up to 60 mm:<br>0,9 $\mu\text{m}$<br><br>for length up to 271 mm:<br>1,1 $\mu\text{m}$ |         |
|                                                                                                                                                                                                |                                                                                                               | Determination of probing<br>error $P_F$ on a reference<br>sphere according to<br>VDI/VDE 2630 part 1.3:2011                                          | 0,26 $\mu\text{m}$                                                                                |         |
|                                                                                                                                                                                                |                                                                                                               | Determination of probing<br>error $P_S$ on a reference<br>sphere according to<br>VDI/VDE 2630 part 1.3:2011                                          | 0,28 $\mu\text{m}$                                                                                |         |

Abbreviations used:

|       |                                                                                                  |
|-------|--------------------------------------------------------------------------------------------------|
| CMC   | Calibration and measurement capabilities                                                         |
| DIN   | Deutsches Institut für Normung e.V.                                                              |
| DKD-R | guideline of Deutscher Kalibrierdienst (DKD), published by Physikalisch-Technische Bundesanstalt |
| VDE   | Verband der Elektrotechnik, Elektronik und Informationstechnik e.V.                              |
| VDI   | Verein Deutscher Ingenieure e.V.                                                                 |

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.